The Mathematics of Diagnostics Tomography

LauraLynne Duffy
Faculty Mentor: Dr. David Seppala-Holtzman
Department: Mathematics
SJC First Annual Research Symposium
April 2013

What is Tomography?

Tomography is a "radiologic technique for obtaining clear X-ray images of deep internal structures by focusing on a specific plane within the body."

What is Computed Tomography?

- Simply refers to tomography done using data collected from X-rays and a computer.
- All X-ray imaging -such as Computed Tomography- is based on the absorption of X-rays as they pass through the different parts of a patient's body.
- Based on the variable absorption of X-rays by different tissues, computed tomography (CT) imaging, also known as "CAT scanning" (Computerized Axial Tomography), provides a form of imaging known as cross-sectional imaging.

(US Food and Drug Administration)

What is Computed Tomography?

 We can think of cross-sectional images or "slices" of anatomy like the slices in a loaf of bread.



Conventional X-Rays




(www.safetynewsalert.com)



(Percision-Upper Cervical Center of New Jersey)

For a tomograph, or a cross section, the beams are projected in the same plane as the picture.




AIM:

To understand the mathematics behind how a cross section of the human body is reconstructed on a computer screen from many individual photon beam measurements.

How does a CT scan work?

First, we divide the field of view in which the cross section is located into N pixels (picture elements) numbered 1 through N.




X-ray source

How does a CT scan work?

- Divide field into N-pixels
- 2. Send X-rays
 through cross
 section to X-ray
 detector
- 3. Determine X-ray density of each pixel

X-ray source



Modes of Scanning:




FIGURE 1-11 Four basic scanning methods or systems: A, first generation; B, second generation; C, third generation; D, fourth generation.

The Three Basic Steps of a CT Scan:

- 1. The field of view containing the cross section is scanned with X-rays.
- 2. The data from the X-rays is used to compute the X-ray density of each of the pixels in the field of view.
- The pixels are colored shades of gray proportional to their X-ray density, producing the cross sectional image on the video monitor.

AIM (simplified):

To determine the X-ray density of each pixel.

What *type* of mathematics is involved?

The construction of a cross section requires the solution of a large linear system of equations.

Certain algorithms, called algebraic reconstruction techniques (ARTs) can be used to solve these linear systems whose solutions yield the cross sections in digital form.

i.e. the cross section that appears on the computer screen is **the solution to a linear system of equations**.

A <u>linear equation</u> is an equation in which all variables have exponent equal to one.

For example,

$$3x - 2y = 7$$

A general linear equation with *n* variables is written as the following:

$$a x_1 + a x_2 + ... + a x_n = b$$

A solution to a linear equation is a set of n real numbers:

such that when $_{i}^{S}$ is substituted for $_{i}^{X}$ the equation is valid.

A <u>system of linear equations</u> - or simply <u>a linear system</u> - is a set of linear equations, **all with the same variables**.

$$a_{1} = b_{1}$$

$$a_{2} = b_{1}$$

$$a_{2} = b_{1}$$

$$a_{2} = b_{2}$$

$$\vdots$$

$$\vdots$$

$$a_{m,1} = b_{m}$$

$$x_{1} + a_{m,2} + ... + a_{m,m} = b_{2}$$

$$\vdots$$

$$\vdots$$


$$a_{m,1} = b_{m}$$

A general linear system of m equations and n variables is written as follows:

m equations

$$a_{1,1} x_{1} + a_{1,2} x_{2} + ... + a_{1,n} x_{n} = b_{1,n} x_{n} = b_{1,n} x_{n} + a_{2,2} x_{2} + ... + a_{2,n} x_{n} = b_{2,n} x_{n} =$$

A general linear system of m equations and n variables is written as follows:



Once more: a linear system of equations is a set of equations with all the same variables.

$$a_{1} = b_{1} + a_{1} = b_{1}$$

$$a_{2} = b_{1} + a_{2} = b_{2}$$


$$a_{2} = b_{2} + ... + a_{2} = b_{2}$$

$$\vdots$$

$$a_{m,1} = b_{m,2} + ... + a_{m,m} = b_{m,m}$$

The Solution to a Linear System of Equations

An ordered sequence of numbers ${}^{s}_{1, 2, n}$ ${}^{s}_{2, n}$ is a solution to a linear system <u>if it is a solution to each and every</u> <u>equation in the system, simultaneously.</u>



(Anton and Rorres)

A Note About Solutions

The solution to a linear system of equations can either be



-meaning it has exactly one or infinite solutions

inconsistent

-meaning it does not have solution, but an approximate solution can be found

This diagram shows a single pixel with an X-ray beam of roughly the same width as the pixel passing squarely through it.

- The photons constituting the X-ray beam are absorbed by the tissue within the pixel at a rate proportional to the X-ray density of the tissue.
- Quantitative, the X-ray density of the jth pixel as denoted by j and is defined by

$$x_{j} = \ln\left(\frac{photons\ entering\ the\ jth\ pixel}{photons\ leaving\ the\ jth\ pixel}\right)$$

- If the X-ray beam passes through an entire row of pixels, then the number of photons leaving one pixel is equal to the number of photons entering the next.
- If the pixels are numbered 1, 2, ...n, then the additive property of the logarithmic function gives us:

$$x_1 + x_2 + ... + x_n = \ln\left(\frac{photons\ entering\ the\ jth\ pixel}{photons\ leaving\ the\ jth\ pixel}\right)$$

- Except that instead of photons entering and leaving the jth pixel, it is the first pixel of the row over the number of photons leaving the nth pixel of the row.

 Thus, due to the additive property of the logarithmic function, to determine the total X-ray density of a row of pixels, we simply sum the individual pixel densities.

- Next we consider the beam density of the ith beam scan
- Beam density is denoted by bi and represented by

obtained from clinical calibration

 For each beam that passes squarely through a row of n pixels

```
fraction of photons of the
beam that pass through the
row of pixels without being
absorbed

fraction of photons of
the beam that pass
through the cross
section without being
absorbed
```

 Thus, for each jth beam we get the following equation:

$$x_{j_{1}} + x_{j_{2}} + \dots + x_{j_{i}} = b_{i}$$

where bi is known from the clinical and calibration measurements, and x1, x2, ...xn are unknown pixel densities that must be determined.

If we set

$$a_{i,j} = \begin{cases} 1, & \text{if } j = j_1, j_2, \dots j_i \\ 0, & \text{otherwise} \end{cases}$$

In other
words, if the
beam
passes
squarely
through the
pixel

then we can rewrite the equation as the following:

$$a_{1,1}x + a_{1,2}x + ... + a_{1,N}x = b_{1,N}$$

Each 1 is on the left, can be replaced with i, giving us the ith beam equation.

- Either a 1 or a 0 is not the best way to represent the coefficient a, however.
- This is because the beams of a scan do not either squarely pass through a pixel or entirely miss it.
- Instead, a typical beam passes diagonally through each pixel in its path.
- We have three ways to determine a; each one with increasing accuracy.

Determining $a_{i,j}$

$$a = \left(egin{array}{ll} 1, & ext{if the ith beam} \ ext{passes through} \end{array}
ight) \ i,j & ext{center} \ 0, & ext{otherwise} \end{array}$$

 Using any of the three methods to define the a's in the ith beam equation, we can write a set of M beam equations in a complete scan as

$$a_{1,1} x + a_{1,2} x + ... + a_{1,N} x = b_{1,N} N N$$

$$a_{2,1} x + a_{2,2} x + ... + a_{2,N} x = b_{2,N} N N$$

$$\vdots$$


$$a_{M,1} x + a_{M,2} x + ... + a_{M,N} x = b_{M}$$

A linear system of equations!

- In this way we have a linear system of M equations (the M beam equations) in N unknowns (the N pixel densities that field of view containing the cross section was divided into)
- The solution to this linear system of equations -determined using Algebraic Reconstruction Techniques (ARTs)- gives us the X-ray densities of each of N pixels.
- The pixels are then shaded gray proportional to their X-ray densities.

 (Anton and Rorres)

And we get a cross section, or a tomograph.



Works Cited

Anton, Howard, and Chris Rorres. *Elementary Linear Algebra: Applications Version*. 10th ed. Hoboken, N.J.: Wiley, 2010. Print.

British Journal of Cancer. *Outpatient weekly neoadjuvant chemotherapy* followed by radiotherapy for advanced nasopharyngeal carcinoma: high complete response and low toxicity rates. N.d. British Journal of Cancer, UK. *Nature Publishing Group*. Web. 13 Apr. 2010.

capone.mtsu.edu. *Modes of Scanning*. N.d. Modes of Scanning, Middle Tennessee State University. *capone.mtsu.edu*. Web. 10 Apr. 2013.

Encyclopedia Britannica. "tomography (radiology) -- Encyclopedia Britannica." *Encyclopedia Britannica*. N.p., n.d. Web. 10 Apr. 2013.

http://www.britannica.com/FBchecked/topic/598996/tomography>

Precision - Upper Cervical Center of New Jersey. *X-Ray of Skull*. 2013. Digital X-Ray, New Jersey, USA. *Precision - Upper Cervical Center of New Jersey*. Web. 10 Apr. 2013.

US Food and Drug Administration. "What is Computed Tomography?." *US Food and Drug Administration Home Page*. N.p., n.d. Web. 10 Apr. 2013. http://www.fda.gov/radiation-emittingproducts/radiationemittingproductsandprocedures/medicalimaging/medicalx-rays/ucm115318.htm.

www.ct-scan-info.com. *Computed X-Ray Tomography (CT) Scanner*. N.d. CT Scan Information , Vidalink, LLC. *www.ct-scan-info.com*. Web. 10 Apr. 2013.

www.physicscentral.com. *The Geometry of the CT Scan*. N.d. CT Scans, American Physical Society . *Physics Central - Learn How Your World Works*.

Wah 10 Apr 2013

www.safetynewsalert.com. "Worker breaks leg while turning around, gets comp - Safety News Alert." *Safety News Alert - OSHA and safety news for workplace safety professionals.*. N.p., n.d. Web. 10 Apr. 2013.

http://www.safetynewsalert.com/worker-breaks-leg-while-turning-around-gets-comp/>.

Thank you for listening.

Questions?